

Certificate of Analysis

Number: FAPON-FR-T0-017

Edition: B.2

Product overview			
Product Name	AMP-BSA antigen	Catalog Number	FPZ0960
Date of Manufacture	Specific date	Lot Number	Specific lot number
Immunogen	N/A		
Species	N/A	Expressed Host	N/A
Tag	N/A	Clone number	N/A
Isotype	N/A	Molecular weight	N/A
Product buffer solution	10mM PB+150mM NaCl+0.1% ProClin300,pH7.4	Preservative	ProClin300
Storage Conditions	-20±5°C, Avoid multiple freeze/thaw cycles	Date of Expiration	Shelf life:24 months
Applications	Capture	Platforms tested	N/A
Parameters tested on each lot			
Product appearance	Liquid		
Conc.determined by	N/A	Concentration /Recommended titer	Specific Concentration
Purification	N/A	Purity	N/A
Performance:			
Items	Passing criteria	Qualified or not	
Immunoreactivity	In testing the internal control samples, the deviation of the colorimetric level in the test group from the control group is within ±0.5.	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	
Clinical specificity	The 10 negative urine samples tested true negative.	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	
Product stability	N/A	Yes <input type="checkbox"/> No <input type="checkbox"/>	
Miscellaneous	N/A	Yes <input type="checkbox"/> No <input type="checkbox"/>	
Remark	N/A		
Conclusion	Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>		
Quality signature	N/A	Quality inspection date	N/A

管理者代表: 
Management representative

广东省东莞市松山湖台湾高科技园花露路5号
No.5 Huailan Road, Taiwan High-tech industrial park, Songshan Lake, Dongguan, Guangdong, 523808, China
0769-22898886 marketcn@fapon.com (国内/Domestic) market@fapon.com (国际/International) www.faponbiotech.com