

## Certificate of Analysis

Number: FAPON-FR-T0-017

Edition: B.2

Product overview			
<b>Product Name</b>	Anti Influenza A monoclonal antibody	<b>Catalog Number</b>	BRJINFS101
<b>Date of Manufacture</b>	Specific date	<b>Lot Number</b>	Specific lot number
<b>Immunogen</b>	Recombinant Influenza A virus Nucleoprotein		
<b>Species</b>	Mouse	<b>Expressed Host</b>	CHO
<b>Tag</b>	N/A	<b>Clone number</b>	N/A
<b>Isotype</b>	N/A	<b>Molecular weight</b>	150kDa
<b>Product buffer solution</b>	10mM PB+150mM Nacl+0.1% ProClin300,pH7.4	<b>Preservative</b>	ProClin300
<b>Storage Conditions</b>	2~8°C	<b>Date of Expiration</b>	Shelf life:36 months
<b>Applications</b>	Detection	<b>Platforms tested</b>	GICA
Parameters tested on each lot			
<b>Product appearance</b>	Liquid		
<b>Conc.determined by</b>	UV absorption	<b>Concentration /Recommended titer</b>	Specific Concentration
<b>Purification</b>	Protein A chromatography+Ion-exchange chromatography	<b>Purity</b>	≥90%
<b>Performance:</b>			
<b>Items</b>	<b>Passing criteria</b>		<b>Qualified or not</b>
<b>Immunoreactivity</b>	In testing the internal control samples, the deviation of the colorimetric level in the test group from the control group is within ± 0.5.		<b>Yes</b> ✓ <b>No</b> □
<b>Clinical specificity</b>	The results are negative in testing 1x PBS buffer, deionized water, physiological saline and 50 mM carbonate buffer (pH 9.51).		<b>Yes</b> ✓ <b>No</b> □
<b>Product stability</b>	N/A		<b>Yes</b> □ <b>No</b> □
<b>Miscellaneous</b>	N/A		<b>Yes</b> □ <b>No</b> □
<b>Remark</b>	N/A		
<b>Conclusion</b>	<b>Pass</b> ✓ <b>Fail</b> □		

Quality signature	N/A	Quality inspection date	N/A
-------------------	-----	-------------------------	-----

管理者代表:   
Management representative