

Certificate of Analysis

Number: FAPON-FR-T0-017

Edition: B.2

| Product overview | | | | | |
|-------------------------------|--|-------------------------------------|------------------------|------------------|--|
| Product Name | Anti D-Dimer recombinant antibody | Catalog Number | BRNDIMS103 | | |
| Date of Manufacture | Specific date | Lot Number | Specific lot number | | |
| Immunogen | D-dimer | | • | | |
| Species | Mouse | Expressed Host | СНО | | |
| Тад | N/A | Clone number | 32D2 | | |
| lsotype | lgG2b | Molecular weight | 150 kDa | | |
| Product buffer solution | 10mM PB+150mM Nacl+0.1% ProClin300,pH7.4 | Preservative | ProClin300 | | |
| Storage Conditions | -20±5°C, Avoid multiple freeze/thaw cycles | Date of Expiration | Shelf life:36 months | | |
| Applications | Capture | Platforms tested | ELISA、CLIA、FIA、GICA | | |
| Parameters tested on each lot | | | | | |
| Product appearance | Liquid | | | | |
| Conc.determined by | UV absorption | Concentration /Recommended titer | Specific Concentration | | |
| Purification | N/A | Purity | ≥90% | | |
| Performance: | | | | | |
| Items | Passing criteria | | | Qualified or not | |
| Immunoreactivity | At the same working concentration, the deviation of the activity in the test group from the control group is within \pm 15%. | | | Yes√ No□ | |
| Clinical specificity | N/A | | | Yes□ No□ | |
| Product stability | N/A | | | Yes□ No□ | |
| Miscellaneous | N/A | | | Yes□ No□ | |
| Remark | N/A | | | | |
| Conclusion | Pass√ Fail □ | | | | |
| Quality signature | N/A | Quality inspection date | N/A | | |

管理者代表: 7 、 4 、 VA

○ 广东省东莞市枢山湖台湾高科技田花道路5号 No.5 Hualian Road, Taiwan High-tech industrial park, Songshan Lake, Dongguan, Guangdong, 523808, China
↓ 0769-22898885
▲ marketcn@fapon.com(国内/Domestic)_market@fapon.com(国际/International)

